

QC Certificate

Product Name	SureSeq Myeloid Plus Panel
Product Code	770002
Lot Number	227842520

Acceptance Criteria

Number of Lanes: 16
Starting Input DNA: 500 ng

SureSeq Pan Myeloid Panel	Specification	Pass/Fail
Mean Target Coverage:	≥700x	Pass
100% Target Bases Covered at:	≥30x	Pass
% Reads Covered at >20% Mean:	≥99%	Pass

Name	Position	Date
Kaajal Reeves	QC Manager	30 June 2022

This product has been released by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

OGT

T: +44(0)1865 856800 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com

Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.