

Quality Control Test Certificate

Product Name	SureSeq NGS Index Kit - Collection B (48)
Product No.	500072
Batch No.	5316270-2
Certificate Issue No.	1

Acceptance Criteria

All Indexes of the SureSeq NGS Index Kit were tested simultaneously using a custom developed protocol. An artificial reference genome was run as an equivalently balanced load on an Illumina Sequencing Platform, and verified to meet the below series of quality specifications:

	Specification	Pass/Fail
Minimum Read Depth:	>420,000x	PASS
Single Index Cross Contamination:	<0.75%	PASS
Index Purity:	≥99.2%	PASS
Sample Purity:	≥99.2%	PASS

This product has been approved for release by:

Name: Vincent Wong Position: Quality Associate Date: 18 September 2018

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com
 Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.