

## Quality Control Test Certificate

Product name	SureSeq MyPanel Custom Core MPN - 5 Exon Panel
Product no.	602005
Batch no.	2963942
Certificate issue no.	1

### Acceptance Criteria

Number of samples per MiSeq lane: 16

DNA input: 1000 ng

SureSeq MyPanel Custom Core MPN - 5 Exon Panel	Specification	Pass/Fail
Mean Target Coverage:	>3000x	PASS
100% Target Bases Covered at:	≥1000x	PASS
% Target Bases Covered at >20% Mean:	99.0%	PASS

This product has been approved for release by:

Name: Vincent Wong      Position: Quality Associate      Date: 21 January 2019

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

### Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com

Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.