

QC Certificate

Product Name	CytoSure HT Genomic DNA Labelling Kit (96)
Product Code	500040
Lot Number	17578032

Acceptance Criteria

Test	Specification	Pass/Fail	
Dye incorporation	Cy3	>250 pmol	Pass
	Cy5	>250 pmol	Pass
DNA yield	>6µg	Pass	
Array performance (Derivative Log Ratio Spread)	<0.25	Pass	

Name	Position	Date
Kaajal Reeves	QC Manager	21 April 2022

This product has been released by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

OGT

T: +44(0)1865 856800 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com

Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.