



A Sysmex Group Company



Instructions For Use

REF: USA-LPH022

CBFβ (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit



PROFESSIONAL USE ONLY

Further information available at www.ogt.com

Intended Use

The CBFβ (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit is a fluorescence in situ hybridization (FISH) Test used to detect rearrangement of the chromosome 16 causing the CBFβ-MYH11 (CBFB-MYH11) fusion in fixed bone marrow specimens from patients with acute myeloid leukemia (AML). The test is indicated for characterization of patient specimens consistent with World Health Organization (WHO) guidelines for Classification of Tumours of Haematopoietic and Lymphoid Tissues (Revised 4th Edition) and in conjunction with other clinicopathological criteria. The assay results are intended to be interpreted by a qualified pathologist or cytogeneticist. The test is not intended for use as a stand-alone diagnostic, disease screening, or as a companion diagnostic

Principles of the test


Fluorescence *in situ* hybridization (FISH) is a technique that allows DNA sequences to be detected on metaphase chromosomes or in interphase nuclei from fixed cytogenetic samples. The technique uses DNA probes that hybridize to entire chromosomes or single unique sequences and serves as a powerful adjunct to G-banded cytogenetic analysis. This technique can now be applied as an essential investigative tool within prenatal, hematological and solid tumor chromosomal analysis. Target DNA, after fixation and denaturation, is available for annealing to a similarly denatured, fluorescently labelled DNA probe, which has a complementary sequence. Following hybridization, unbound and non-specifically bound DNA probe is removed and the DNA is counterstained for visualization. Fluorescence microscopy then allows the visualization of the hybridized probe on the target material.

Probe Information

CBFβ (CBFB) /MYH11 Translocation, Dual Fusion FISH Probe Kit: The CBFB (core-binding factor beta subunit) gene is located at 16q22, whilst the MYH11 (myosin heavy chain 11) gene is located at 16p13.1. The inversion inv(16)(p13.1;q22.1) and the translocation t(16;16)(p13.1;q22.1) give rise to the CBFB-MYH11 fusion gene. Acute myeloid leukemias with inv(16)(p13.1;q22.1) or t(16;16)(p13.1;q22.1) form a recognized disease entity according to the World Health Organization (WHO) classification of myeloid neoplasms and acute leukemia. These rearrangements are frequently found in patients with a myelomonocytic subtype with increased bone marrow eosinophils, AML FAB (French-American-British classification) type M4Eo. Cases of therapy-related AML may also have this rearrangement.

Probe Specification

CBFβ, 16q22, Texas Red
MYH11, 16p13.1, FITC green



The CBFβ (CBFB) /MYH11 Translocation, Dual Fusion FISH Probe Kit consists of a 617kb probe, labeled in Texas Red, covering a region, within 16q22 including the *CBFB* gene; and a 621kb probe, labeled in FITC green, covering a region within 16p13.1 including the *MYH11* gene.

Materials Provided

Probe: 100µl per vial (LPH022: 10 tests)

Amount of Texas Red CBFβ probe: 14.0-23.6 ng/test

Amount of FITC green MYH11 probe: 54.8-82.0 ng/test

The probes are provided premixed in hybridization solution (formamide; dextran sulfate; saline-sodium citrate (SSC)) and are ready to use.

Counterstain: 150µl per vial (DES150L: 15 tests)

The counterstain is DAPI antifade (0.125µg/ml DAPI (4,6-diamidino-2-phenylindole)).

Warnings and Precautions

1. For *in vitro* diagnostic use. For professional use only.
2. For prescription use only
3. Wear gloves when handling DNA probes and DAPI counterstain.
4. Probe mixtures contain formamide, which is a teratogen; do not breathe fumes or allow skin contact. Wear gloves, a lab coat, and handle in a fume hood. Upon disposal, flush with a large volume of water.
5. DAPI is a potential carcinogen. Handle with care; wear gloves and a lab coat. Upon disposal, flush with a large volume of water.
6. Dispose of all hazardous materials according to your institution's guidelines for hazardous waste disposal.
7. Operators must be capable of distinguishing the colors red, blue and green.
8. The probe is intended to be used only at the concentration provided and is not intended to be diluted
9. Failure to adhere to the protocol and the reagents may affect the performance and lead to false positive/negative results.

Labelling according to GHS-US hazard label requirements

Hazard pictograms (GHS-US):



GHS07



GHS08

Signal word (GHS-US): Danger

Hazardous ingredients: Formamide <100%

Hazard statements (GHS-US):

H315 - Causes skin irritation

H319 - Causes serious eye irritation

H360 - May damage fertility or the unborn child

Precautionary statements (GHS-US):

P202 - Do not handle until all safety precautions have been read and understood

P280 - Wear eye protection, protective clothing, protective gloves

P302+P352 - IF ON SKIN: Wash with plenty of soap and water

P305+P351+P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing

P308+P313 - IF exposed or concerned: Get medical advice/attention

P362+P364 - Take off contaminated clothing and wash it before reuse
P501 - Dispose of contents/container to hazardous or special waste collection point, in accordance with local, regional, national and/or international regulation

Refer to the Safety Data Sheet for more information.

Storage and Handling

Store the kit at -20°C until the expiry date indicated on the kit label. Store the probe and counterstain vials in the dark. Ensure that exposure of the probe and counterstain to laboratory lights is limited at all times.

Equipment and Materials Necessary but not Supplied

1. Hotplate (with a solid plate and accurate temperature control up to 80°C).
2. Variable volume micropipettes and tips range 1µl - 200µl.
3. Water bath with accurate temperature control at 37°C and 72°C.
4. Microcentrifuge tubes (0.5ml)
5. Centrifuge.
6. Fluorescence microscope (Please see Fluorescence Microscope Recommendation section).
7. Plastic or heat-resistant glass Coplin jars.
8. Forceps.
9. pH meter (or pH indicator strips capable of measuring pH 6.5 – 8.0).
10. Humidified container.
11. Fluorescence grade microscope lens immersion oil.
12. Bench top centrifuge.
13. Microscope slides.
14. 24x24mm coverslips or equivalent.
15. Timer.
16. 37°C incubator.
17. Rubber solution glue.
18. Vortex mixer.
19. Graduated cylinders.
20. Magnetic stirrer.
21. Calibrated thermometer.
22. Clear nail varnish

Reagents Needed but not Supplied

1. 20x saline-sodium citrate (SSC) Solution.
2. 100% Ethanol.
3. Tween-20.
4. 1M Sodium hydroxide (NaOH)
5. 1M Hydrochloric acid (HCl).
6. Purified water.
7. Immersion oil.

Fluorescence Microscope Recommendation

Use a 100-watt mercury lamp, or equivalent, and oil immersion plan apochromat objectives x63 or x100 for optimal visualization. Use a triple bandpass filter DAPI/FITC/Texas Red or a dual bandpass filter FITC/Texas Red for optimal simultaneous visualization of the green and red fluorophores. The aqua fluorophore has specificity to the aqua and DEAC spectrum (single bandpass aqua or DEAC filter is required).

Check the fluorescence microscope before use to ensure it is operating correctly. Use immersion oil that is suitable for fluorescence microscopy and formulated for low autofluorescence. Avoid mixing DAPI Antifade with microscope immersion oil as this will obscure signals. Follow manufacturers' recommendations with regard to the life of the lamp and the age of the filters.

Sample Preparation

The kit is designed for use on bone marrow cells fixed in Carnoy's solution (3:1 methanol/acetic acid) fixative, that are prepared according to the laboratory or institution guidelines. Prepare air dried samples on microscope slides according to standard cytogenetic procedures. The AGT *Cytogenetics Laboratory Manual* contains recommendations for specimen collection, culturing, harvesting and for slide making.

Solution Preparation

Ethanol Solutions

Dilute 100% ethanol with purified water using the following ratios and mix thoroughly.

70% Ethanol - 7 parts 100% ethanol to 3 parts purified water.

85% Ethanol - 8.5 parts 100% ethanol to 1.5 parts purified water.

Store the solutions for up to 6 months at room temperature.

2xSSC Solution

Dilute 1 part 20xSSC Solution with 9 parts purified water and mix thoroughly. Check pH and adjust to pH 7.0 using NaOH or HCl as required. Store the solution for up to 4 weeks at room temperature.

0.4xSSC Solution

Dilute 1 part 20xSSC Solution with 49 parts purified water and mix thoroughly. Check pH and adjust to pH 7.0 using NaOH or HCl as required. Store the solution for up to 4 weeks at room temperature.

2xSSC, 0.05% Tween-20

Dilute 1 part 20xSSC Solution with 9 parts purified water. Add 5µl of Tween-20 per 10ml of diluted SSC solution and mix thoroughly. Check pH and adjust to pH 7.0 using NaOH or HCl as required. Store the solution for up to 4 weeks at room temperature.

FISH Protocol

(Note: Ensure that exposure of the probe and counterstain to laboratory lights is limited at all times).

Slide preparation

1. Spot the cell sample onto a glass microscope slide. Allow to dry.

2. Immerse the slide in 2xSSC for 2 minutes at room temperature (RT) without agitation.
3. Dehydrate in an ethanol series (70%, 85% and 100%), each for 2 minutes at RT.
4. Allow to dry.

Pre-Denaturation

5. Remove the probe from the freezer and allow it to warm to room temperature (RT). Briefly centrifuge tubes before use.
6. Ensure that the probe solution is uniformly mixed with a pipette or a vortex mixer.
7. Remove 10µl of probe per test and transfer it to a microcentrifuge tube. Quickly return the remaining probe to -20°C.
8. Place the probe and the sample slide to prewarm on a 37°C (+/- 1°C) hotplate for 5 minutes.
9. Spot 10µl of probe mixture onto the cell sample and carefully apply a coverslip. Seal with rubber solution glue and allow the glue to dry completely.

Denaturation

10. Denature the sample and probe simultaneously by heating the slide on a hotplate at 75°C (+/- 1°C) for 2 minutes.

Hybridization

11. Place the slide in a humid, lightproof container at 37°C (+/- 1°C) overnight.

Post-Hybridization Washes

12. Remove the DAPI from the freezer and allow it to warm to room temperature (RT).
13. Remove the coverslip and all traces of glue carefully.
14. Immerse the slide in 0.4xSSC (pH 7.0) at 72°C (+/- 1°C) for 2 minutes without agitation.
15. Drain the slide and immerse it in 2xSSC, 0.05% Tween-20 at RT (pH 7.0) for 30 seconds without agitation.
16. Drain the slide and apply 10µl of DAPI antifade onto each sample.
17. Cover with a 24x24mm coverslip, remove any bubbles.
18. Edge the slide with clear nail varnish to seal.
19. Allow the color to develop in the dark for 10 minutes.
20. View with a fluorescence microscope. (See **Fluorescence Microscope Recommendation**.)

Stability of Finished Slides

Hybridized slides remain analyzable for up to 1 month if stored in the dark at refrigerator temperature.

Procedural Recommendations

1. Baking or ageing of slides may reduce signal fluorescence
2. Hybridization conditions may be adversely affected by the use of reagents other than those provided or recommended by CytoCell Ltd.
3. Use a calibrated thermometer for measuring temperatures of solutions, waterbaths and incubators as these temperatures are critical for optimum product performance.
4. The wash concentrations, pH and temperatures are important as low stringency can result in non-specific binding of the probe and too high stringency can result in a lack of signal.
5. Incomplete denaturation can result in lack of signal and over denaturation can also result in non-specific binding.
6. Over hybridization can result in additional or unexpected signals.
7. Users should optimize the protocol for their own samples prior to using the test for diagnostic purposes.
8. Suboptimal conditions may result in non-specific binding that may be misinterpreted as a probe signal.

Interpretation of Results





Assessing slide quality

The slide should not be analyzed if:

- Signals are too weak to analyze in single filters - in order to proceed with analysis, signals should appear bright, distinct and easily evaluable,
- There are high numbers of clumped/overlapping cells obstructing the analysis
- >50% of the cells are not hybridized
- There is excess of fluorescent particles between cells and/or a fluorescent haze - in optimal slides the background should appear dark or black and clean.
- Cell nucleus borders cannot be distinguished and are not intact.

Analysis guidelines

- Two analysts should analyze each sample
- Each analyst should score independently 100 nuclei for each sample. The first analyst should start the analysis from the left side of the spot and the second analyst from the right one, so that the analysts are examining different areas of the hybridization and not scoring the same cells.
- Each analyst should document their results in separate sheets.
- Analyze only intact nuclei, not overlapped or crowded nuclei or nuclei covered by cytoplasmic debris or high degree of autofluorescence.
- Avoid areas where there is excess of cytoplasmic debris or non-specific hybridization.
- Signal intensity may vary, even with a single nucleus. In such cases, use single filters and/or adjust the focal plane.
- In suboptimal conditions signals may appear diffuse. If two signals of the same color touch each other, or the distance between them is no greater than one signal width, or when there is a faint strand connecting the two signals, count as one signal
- If in doubt about whether a cell is analyzable or not, then do not analyze it.
- If the abnormal signal count result falls up to 2x the cut-off, the sample should be re-processed and re-analyzed

Analysis guidelines	
	Do not count – nuclei are too close together to determine boundaries
	Do not count overlapping nuclei – all areas of both nuclei are not visible
	Count as two red signals and two green signals – one of the two red signals is diffuse
	Count as two red signals and two green signals – the gap in one red signal is less than two probe widths


Third analyst requirements

Each of the two analysts will analyze 100 cells. In some cases, depending on the number of abnormal nuclei each analyst has seen, a third reader may be required. The rules for a third analyst are detailed in the text and the flowchart below.

A third analyst is required if:


- One of the analyst's abnormal signal count is above the cut-off and the other below or at the cut-off.
- The two-analysts' abnormal counts differ by more than 5 and either analyst's abnormal signal count is equal to or lower than twice the cut-off.
- The two analysts' abnormal signal patterns differ by more than 15.

The third analyst will also analyze 100 nuclei. Of the three scores obtained, the two that are closer to each other should be used. If the 3 scores are equidistant, then the median value should be doubled and used.




Expected Results

Expected normal signal pattern



In a normal cell, two red and two green signals (2R, 2G) are expected.

Expected abnormal signal pattern



In a cell with an inv(16)(p13;q22) or a t(16;16)(p13;q22) the expected signal pattern will be one red, one green and two fusions (1R, 1G, 2F).

Other signal patterns are possible in aneuploid/unbalanced specimens.

Known Cross-Reactivity

No known cross reactivity.

Limitations of the procedure

Reporting and interpretation of FISH results should be consistent with professional standards of practice and should take into consideration other clinical and diagnostic information. This kit is intended as an adjunct to other diagnostic laboratory tests and therapeutic action should not be initiated on the basis of the FISH result alone. Failure to adhere to the protocol may affect the performance and lead to false results.

Each lab is responsible for establishing their own cut-off values. Each laboratory should test sufficiently large number of samples to establish normal population distribution of the signal levels and to assign a cut-off value. The product is for professional use only and is intended to be interpreted by a qualified Pathologist or Cytogeneticist.

The device has not been specifically validated in patients with <20% blast count.

Specific Performance Characteristics

Analytical Specificity

Analytical specificity is the percentage of signals that hybridize to the correct locus and no other location. The analytical specificity of the CBFβ (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit was established by analyzing a total of 200 target loci from metaphase chromosomes prepared from five normal male peripheral blood samples. The analytical specificity was calculated as the number of FISH signals that hybridized to the correct locus divided by the total number of FISH signals hybridized. The result was then multiplied by 100, to give a percentage, and is given with a 95% confidence interval. The analytical specificity of the CBFβ (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit probes was 100%, as shown in Table 1.

Table 1. Analytical Specificity for the CBFβ (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit

Probe	Target	Number of metaphase chromosomes hybridized	Number of correct hybridized loci	Analytical Specificity (%)	95% Confidence Interval (%)
CBFβ, Red	16q22	200	200	100	98.12 - 100
MYH11, Green	16p13.1	200	200	100	98.12 - 100

Analytical Sensitivity

Analytical sensitivity is defined as the percentage of scoreable interphase cells with the expected normal signal pattern.

The analytical sensitivity of the CBFβ (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit, was established by analyzing interphase nuclei from 25 bone marrow samples, selected from the intended population for the probe. Each sample was analyzed by two independent analysts and the signal pattern of each interphase was recorded. Each analyst analyzed 100 nuclei per sample, for a total of 200 nuclei per sample, resulting in 5000 scoreable nuclei per probe evaluated.

The sensitivity was calculated as the percentage of scoreable cells with the expected signal pattern and is given below with a 95% confidence interval.

Table 2. Analytical Sensitivity for the CytoCell CBFβ (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit

Probe	Number of interphase nuclei with the expected normal signal pattern	Total number of interphase nuclei analyzed	Analytical sensitivity (%)	95% Confidence interval (%)
USA-LPH022 CBFβ (CBFB)/MYH11 Translocation, Dual Fusion Probe	4947	5000	98.94	98.62 - 99.19

Characterization of Normal Cut-off Values

The limit of detection of a FISH test is established by calculating the upper limit of the abnormal signal pattern in normal cells. This upper limit constitutes the "normal cut-off value". The normal cut-off value, in association with FISH probes, is the maximum percentage of scoreable interphase cells with a specific abnormal signal pattern at which a sample is considered normal for that signal pattern.

An analytical result above the normal cut-off (upper reference limit) is deemed to have detected the rearrangement. Cut-off values are given for each expected abnormal signal pattern. Conversely, an analytical result below the cut-off is deemed to be negative for the rearrangement.

The normal cut-off was calculated using the Microsoft Excel β inverse function (BETA.INV). Beta inverse uses the upper bound of a one-sided 95% confidence interval of the binomial distribution rearrangement.

Laboratories must verify cut-off values using their own data.

The normal cut-off for the CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit, was established, using the data from 1300 bone marrow samples negative for the rearrangement the probe is intended to detect. Two independent analysts analyzed a target number of 100 interphase nuclei per sample.

The normal cut-off value for the '1R, 1G, 2F' signal pattern when using the CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit was calculated to be 2.3% using the beta inverse function.

Table 3. Characterization of Normal Cut-off Values for the CytoCell CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit

Abnormal signal pattern	Number of samples analyzed to generate the cut-off	Number of nuclei evaluated per sample	Maximum number of false positive signal pattern	Normal cut-off value (per 200 nuclei)	Normal cut-off value (%)
1R, 1G, 2F	1300	200	1	5	2.3

Reproducibility

Reproducibility is a measure of the variability of a test and has been established in terms of site-to-site, day-to-day and batch-to-batch variability. Site-to-site and day-to-day reproducibility were assessed by analyzing two replicates of two high-positive, two near cutoff/low-positive and two normal bone marrow samples on each of five days at three sites. For each sample, a total of 200 interphase cells was scored by two analysts and the probe signal patterns recorded. The percentage of cells with the expected signal pattern was calculated. Lot-to-lot reproducibility, was assessed by analyzing four replicates of one high-positive, one low-positive and one normal bone marrow sample on three different batches of probe at one site. A total of 200 interphase nuclei were analyzed per spot, by two independent analysts and the probe signal patterns recorded. The results of the data were analyzed and the overall, negative and positive percent agreement for each sample summarized. The agreement percentages for the High Positive and Negative samples were 100% and 100% respectively.

Table 4. Specimen agreement Site-to-Site Reproducibility for the CytoCell CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit

Specimen	Agree	Disagree	Total	Agreement (%)
Negative 1	30	0	30	100
Negative 2	30	0	30	100
Near cut-off 1	8	22	30	27
Near cut-off 2	13	17	30	43
Positive 1	30	0	30	100
Positive 2	30	0	30	100

Table 5. Site-to-Site Analysis of Variance - Standard Deviation (SD) for the CytoCell CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit

Specimen	N	Mean of abnormal percentage	Intra-Day SD	Inter-Day SD	Between-site SD	Total SD
Negative 1	30	0.00	0.00	0.00	0.00	0.00
Negative 2	30	0.02	0.00	0.00	0.00	0.09
Near cut-off 1	30	1.45	0.00	0.51	0.00	1.30
Near cut-off 2	30	2.88	0.75	0.00	0.10	2.47
Positive 1	30	69.67	0.00	0.84	0.64	4.55
Positive 2	30	73.60	0.00	1.76	3.20	7.50

Table 6. Additional supporting data for precision of the CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit

Specimen	Mean of abnormal percentage	Range	Agreement replicates	Agreement (%)	Intra-Day SD	Inter-Day SD	Total SD
Near cut-off 1	4.62	2.5-7	20/20	100	0.56	0.25	1.34
Near cut-off 2	9.75	4.5-13.5	20/20	100	0.00	0.00	2.42

Comparison of Results from Cited Published Literature

The CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit was used by two different clinical laboratories on unselected AML/MDS samples. The number of abnormal samples identified showing the expected abnormal signal pattern was compared with the expected incidence rate from a literature source review. The results fell within the expected range derived from the literature review.

Cited published literature may discuss device uses that have not been approved or cleared by FDA.

Table 7a. Results from Cited Published Literature for the CytoCell CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit – Confirmed AML Cases

Condition	Literature Source 1 Papaemmanuil et al.	Literature Source 2 Grimwade et al.	Literature Source 3 Dores et al.
Was the specific device under review in the submission used in the study?	No	No	No
Was the specimen type in the study representative of the claimed specimen type(s)	Yes	Yes	Yes
Target population (disease status)	Confirmed AML	Confirmed AML	Confirmed AML
Upper reference limit - 'Cut-off value' (percentage and per 200 nuclei)	N/A	N/A	N/A
Total Number of specimens tested for each claimed type	1540	5876	19497
Number of specimens with a positive probe result	N/A	N/A	N/A
Range of positive probe results	N/A	N/A	N/A
Source incidence rate for rearrangement %	5.26%	5%	1.04%

Table 7b. Results from Clinical Laboratory Testing using the CytoCell CBF β (CBFB)/MYH11 Translocation, Dual Fusion FISH Probe Kit

Condition	Data Source 1	Data Source 2
Was the specific device under review in the submission used in the study?	Yes	Yes
Was the specimen type in the study representative of the claimed specimen type(s)	Yes	Yes
Target population (disease status)	Known or suspected MDS or AML	Known or suspected MDS or AML
Upper reference limit - 'Cut-off value' (percentage and per 200 nuclei)	2.3% or 5 1R1G2F patterns per 200 scoreable interphase nuclei	2.3% or 5 1R1G2F patterns per 200 scoreable interphase nuclei
Total Number of specimens tested for each claimed type	100	266
Number of specimens with a positive probe result	2	7
Range of positive probe results	83.5%-93.5%	14%-99.5%
Source incidence rate for rearrangement % (95% CI)	2% (0.24% to 7.04%)	2.63% (1.06% to 5.35%)
Expected range from Literature	1.04% - 5.26%	

Additional Information

For additional product information please contact the CytoCell Technical Support Department.

T: +44 (0)1223 294048

E: techsupport@cytoCELL.com









W: www.cgt.com

References

- Swerdlow et al., (eds.) *WHO Classification of Tumours of Haematopoietic and Lymphoid Tissue*, Lyon, France, 4th edition, IARC, 2017
- Hernandez JM, Gonzalez MB, Granada I, Gutierrez N, Chillon C, Ramos F, et al. *Detection of inv(16) and t(16;16) by fluorescence in situ hybridization in acute myeloid leukemia M4Eo*. Haematologica. Italy; 2000 May;85(5):481-5.
- Arsham, MS., Barch, MJ. and Lawce HJ. (eds.) (2017) *The AGT Cytogenetics Laboratory Manual*. New Jersey: John Wiley & Sons Inc.
- Wiktor AE, Van Dyke DL, Stupca PJ, et al. *Preclinical validation of fluorescence in situ hybridization assays for clinical practice*. Genet Med. 2006;8(1):16-23.
- Wilson, EB. *Probable inference, the law of succession, and statistical inference*. J. Am. Stat. Assoc. 1927;22:209-212.
- Mascarello JT, Hirsch B, Kearney HM, et al. *Section E9 of the American College of Medical Genetics technical standards and guidelines: fluorescence in situ hybridization*. Genet Med. 2011;13(7):667-675.
- Wolff DJ, Bagg A, Cooley LD et al. *Guidance for Fluorescence in Situ Hybridization Testing in Hematologic Disorders*. J Med Diagnost. 2007;9(2):134-143.
- Papaemmanuil E, Gerstung M, Bullinger L, Gaidzik VI, Paschka P, Roberts

ND, et al. *Genomic Classification and Prognosis in Acute Myeloid Leukemia*. N Engl J Med. 2016 Jun 9;374(23):2209–21.

9. Grimwade, D. et al., 2010. *Refinement of cytogenetic classification in acute myeloid leukemia: Determination of prognostic significance of rare recurring chromosomal abnormalities among 5876 youngeradult patients treated in the United Kingdom Medical Research Council trials*. Blood, 116(3), pp.354–365.
10. Does, G.M. et al., 2011. *Acute leukemia incidence and patient survival among children and adults in the United States, 2001-2007*. Blood, 119(1), pp.34–43.

REF	EN: Catalogue number
	EN: <i>In vitro</i> diagnostic device
	EN: Batch code
	EN: Consult instructions for use
	EN: Manufacturer
	EN: Use by
	EN: Temperature limitation
	EN: Sufficient for <n> tests
	EN: Contents
Rx only	EN: For prescription use only

Patents and Trademarks

CytoCell is a registered trademark of CytoCell Ltd.
This product contains technology licensed from Life Technologies Corporation that is available for human diagnostics or life science research use only.



Cytocell Ltd.

Oxford Gene Technology,
418 Cambridge Science Park,
Milton Road,
Cambridge, CB4 0PZ, UK
T: +44(0)1223 294048
F: +44(0)1223 294986
E: probes@cytoCell.com
W: www.ogt.com