

## Quality Control Test Certificate

Product name	SureSeq Blocking Oligos
Product no.	500078
Batch no.	5315471
Certificate issue no.	1

### Acceptance Criteria

*The new stock of SureSeq Blocking Oligos was tested using a SureSeq NGS Cancer Panel for target enrichment with a human genomic DNA library. Current and New stock was compared and run on an Illumina sequencer to verify that it meets the below series of quality specifications.*

SureSeq NGS Cancer Panel	Specification	Pass/Fail
Mean Target Coverage:	>1000x	PASS
100% Target Bases Covered at:	≥1000x (Target Regions) ≥500x (Remainder)	PASS
% Target Bases Covered at >20% Mean:	99.0%	PASS

This product has been approved for release by:

Name: Vincent Wong      Position: Quality Associate      Date: 15 May 2018

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

### Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com  
 Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.