

Quality Control Test Certificate

Product name	SureSeq NGS Library Preparation Kit (16)
Product no.	500074
Batch no.	B057420
Certificate issue no.	1

Acceptance Criteria

The SureSeq NGS LPK (16) was used to create a genomic DNA sequencing library and then sequenced on an Illumina sequencer to verify that it meets the below series of quality specifications:

	Specification	Pass/Fail
% Reads Aligned:	> 90%	PASS
% Duplication:	<10%	PASS
% Reads Mapped in Pairs:	>90%	PASS

This product has been approved for release by:

Name: Vincent Wong Position: Quality Associate Date: 09 May 2018

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com
 Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.